

## Test Certificate

**Reference to Test Report No:** 415.560/3a

**Client:** Semperit Technische Produkte GmbH  
Division Medizintechnik  
Modecenterstraße 22  
1031 Wien

**Order:** Test of medical examination gloves  
for compliance with to ÖNORM EN 455-1, -2, -3

**Samples:** Medical examination gloves  
Semperguard Latex powder-free, innercoated

**Receipt of Samples:** 2015-04-13

**Tests:** The tests were carried out in the individual technical departments within the scope of competence of the authorised signatories according to the OFI QM manual.

### Results:

| Test name                                                                                                                   | Test method           | Date of testing          | Results |
|-----------------------------------------------------------------------------------------------------------------------------|-----------------------|--------------------------|---------|
| Medical gloves for single use -<br>Testing for freedom of holes<br>acc. to ÖNORM EN 455-1                                   | ÖNORM EN 455-1 (2001) | 2015-05-26               | Comply  |
| Medical gloves for single use -<br>Determination of force at break<br>Determination of dimensions<br>acc. to ÖNORM EN 455-2 | ÖNORM EN 455-2 (2011) | 2015-05-13<br>2015-05-26 | Comply  |
| Medical gloves for single use -<br>Determination of removable surface<br>powder acc. to ÖNORM EN 455-3                      | ÖNORM EN 455-3 (2007) | 2015-06-10               | Comply  |

The testing results relate only to the samples analysed and refer to the detailed information of OFI Test report No. 415.560/3a.

Approved: .....  
Mag. Elisabeth Novak, Director in charge



Date: 01.07.2015

This certificate comprises 1 page.

It may be made available to third parties, either free of charge or against payment, if the full wording and if ofi is expressly named as the author. All tests applied are subject to a quality assurance programme according to EN ISO/IEC 17025:2007.

OFI Technologie & Innovation GmbH  
t: +43 1 798 16 01 - 0  
f: +43 1 798 16 01 - 8  
office@ofi.at  
www.ofi.at

Firmensitz:  
1030 Wien, Franz-Grill-Straße 5, Arsenal, Objekt 213  
Weitere Standorte:  
1110 Wien, Brehmstraße 14a  
2700 Wr. Neustadt, Viktor Kaplan Straße 2 / Bauteil C