

Test Certificate

Reference to Test Report No: 415.560/2a

Client: Semperit Technische Produkte GmbH
Division Medizintechnik
Modecenterstraße 22
1031 Wien

Order: Test of medical examination gloves
for compliance with to ÖNORM EN 455-1, -2

Samples: Medical examination gloves
Semperguard, Latex powdered

Receipt of Samples: 2015-04-13

Tests: The tests were carried out in the individual technical departments within the scope of competence of the authorised signatories according to the OFI QM manual.

Results:

Test name	Test method	Date of testing	Results
Medical gloves for single use - Testing for freedom of holes acc. to ÖNORM EN 455-1	ÖNORM EN 455-1 (2001)	2015-04-20	Comply
Medical gloves for single use - Determination of force at break Determination of dimensions acc. to ÖNORM EN 455-2	ÖNORM EN 455-2 (2011)	2015-04-23 2015-04-20	Comply

The testing results relate only to the samples analysed and refer to the detailed information of OFI Test report No. 415.560/2a.

Approved:

Mag. Elisabeth Novak, Director in charge



Date: 01.07.2015

This certificate comprises 1 page.

It may be made available to third parties, either free of charge or against payment, if the full wording and if **ofi** is expressly named as the author. All tests applied are subject to a quality assurance programme according to EN ISO/IEC 17025:2007.

OFI Technologie & Innovation GmbH
t: +43 1 798 16 01 - 0
f: +43 1 798 16 01 - 8
office@ofi.at
www.ofi.at

Firmensitz:
1030 Wien, Franz-Grill-Straße 5, Arsenal, Objekt 213
Weitere Standorte:
1110 Wien, Brehmstraße 14a
2700 Wr. Neustadt, Viktor Kaplan Straße 2 / Bauteil C