

SHIJIAZHUANG HONGRAY GROUP

South Tongda Rd., East Dist. Jinzhou City
Hebei, 052260, China
TEL: 86-311-66179662
www.hongray.com

SPECIFICATION FOR NITRILE EXAMINATION GLOVE 1.0

Product: Nitrile, Powder Free, Blue, 12"length, for examination use, NPF2001-2005 29CM

2.0 Dimensions:

Size	Median Length (mm)	Median Width (mm)	Thickness (mm)(min)	
			Palm	Finger
XS	290	70 ± 10	0.09	0.12
S	290	80 ± 10	0.09	0.12
M	290	95 ± 10	0.09	0.12
L	290	110 ± 10	0.09	0.12
XL	290	120 ± 10	0.09	0.12
Note: Thickness for reference only, not as a basis for rejection.				

3.0 Strength:

Force at break: 6N (Medium)

4.0 Water Leakage Testing: According to ISO2859, G-I, Single Sampling Plan, AQL1.0

5.0 Powder Residues: Not more than 2mg per glove

DOCUMENT NO.: NBR-SETINO-002E/002

INITIALED DATE: 2022.10.08

REVISION DATE: N/A

REVISION TIMES: 0